

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

MANUFACTURER: Shenzhen Yimi Life Technology Co. ,Ltd
305, Building A, Tengbo Industrial Park, Changshangjiang
Street, Longbei Village, Pingshan District, 518118,
Shenzhen, PEOPLE'S REPUBLIC OF CHINA

EUROPEAN REPRESENTATIVE: SHARE INFO CONSULTANT SERVICE LLC REPR SENTANZB RO

PRODUCT/MODEL: Pulse Oximeter/ YM101, YM102, YM103, YM201, YM301
The accessories are used together with the product: Not applicable

UMDNS/GMDN [name/code]: Oximeter / 12853

CLASSIFICATION: Class IIb, Rule 10 According To Annex IX of the MDD

CONFORMITY ASSESSMENT ROUTE: Annex II EXCLUDING (4)

WE, SHENZHEN YIMI LIFE TECHNOLOGY CO. LTD., HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCT(S) MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 INCLUDING AMENDMENTS BY DERECTIVE 2007/47/EC.

ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURER. SHENZHEN YIMI LIFE TECHNOLOGY CO. LTD IS RESPONSIBLE FOT THIS DECLARATION OF CONFORMITY

STANDARDS APPLIED:

NOTIFIED BODY: TÜV SÜD PRODUCT SERVICE GMBH
RIDLERSTR 65, D-80339 M NCHEN, GERMANY

IDENTIFICATION NUMBER: 0123

(EC) CERTIFICATE(S): No.G1 104553 0001 REV.00 VALID UNTIL: 2024-05-26

START OF CE-MARKING: 2020-03-20

PLACE, DATE OF ISSUE: SHENZHEN,

SIGNATURE: YiYAO
NAME YIYAO
GENERAL MANAGER